



3000-IV: Open the instrument

- Switch-off the instrument
- Remove these two screws and open top cover



- Remove these three screws and remove side panel

